

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 09 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130100
A	Name	HUTT, William T.	ID# 3359 Date 11/02/09 ✓
	Agency	AST/Anchor Point Post (ANPE)	Phone # 235-8239 235-3000
	Instrument Location	34115 Sterling Hwy./P.O. Box 007 Anchor Point, AK ✓ 99556	
B	Alco S/N	X124635 ✓	Target Value .080 ✓ High Pressure 1100 psi
	Alco Test Values	.084 ✓ 1 st Alco	.084 ✓ 2 nd Alco
	Signature	William T. Hutt ✓	
(OVER)			


I, Nita J. Bolz, after being first duly sworn, depose (and state as follows) below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130188

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188 ✓

NOVEMBER 02, 2009 ✓

OPERATOR'S NAME: WTHA
HUTT, WILLIAM T
OPERATOR'S NUMBER: 3359
SUBJECT'S LAST NAME: BMA
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI: 11/14/09
ETON

O.L. #: 01234567
DEPT/AGENCY: AST / ANPE
CASE/REPORT: 09-00000
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X124635 ✓

BREATH ANALYSIS

.000 ADJUSTED FOR 30.01 in
ALCO TARGET .000 22:45
BLANK TEST .000 22:46
INTERNAL STANDARD VERIFIED 22:46
ALCO TV 30.01 in .004 22:47
BLANK TEST .000 22:48
SUBJECT SAMPLE WTHA .000 22:48
BLANK TEST .000 22:49
ALCO TV 30.01 in .004 22:50
BLANK TEST .000 22:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130188

NOVEMBER 02, 2009 ✓
TIME 22:40

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 40c
BAROMETER: 30.01 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxy z{|}~■